

## Statement of Accuracy

Device Name: z1HT

Serial: HTE50001

Date: 20/05/2024

**z1HT** data loggers are calibrated for temperature in zLogg's Calibration Chamber.

The reference equipment used is traceable to National Institute of Standards and Technology (NIST).

Temperature Reference Meter: Dostmann P655 Digital Thermometer and can be downloaded here.

This certificate is valid for 2 years from the date of purchase.

Full Range	Temp./Hum. Range	Accuracy
-40°C to +125°C	-40°C to +125°C	±0.2°C